

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102120\
 Data File : VV019055.D
 Acq On : 21 Oct 2020 17:43
 Operator : SY/MD
 Sample : VIBLK63
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK63

Quant Time: Oct 22 07:50:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 22 06:00:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	282181	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	286330	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	137406	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	119525	51.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.92%
7) Chloroethane-d5	1.58	69	106487	57.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	115.72%
11) 1,1-Dichloroethene-d2	2.12	63	196890	44.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.80%
21) 2-Butanone-d5	3.91	46	152133	93.96	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.96%
24) Chloroform-d	4.38	84	224426	53.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.54%
26) 1,2-Dichloroethane-d4	5.05	65	162925	58.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.62%
32) Benzene-d6	5.07	84	459903	55.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.04%
36) 1,2-Dichloropropane-d6	6.09	67	107418	40.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.38%
41) Toluene-d8	7.34	98	331211	44.17	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.34%
43) trans-1,3-Dichloropropene-	7.64	79	58092	44.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.94%
47) 2-Hexanone-d5	8.11	63	81084	81.67	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.67%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	128140	40.91	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.82%
64) 1,2-Dichlorobenzene-d4	11.65	152	140841	50.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.76%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102120\
Data File : VV019055.D
Acq On : 21 Oct 2020 17:43
Operator : SY/MD
Sample : VIBLK63
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK63

Quant Time: Oct 22 07:50:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 22 06:00:52 2020
Response via : Initial Calibration

