

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102120\
 Data File : VV019041.D
 Acq On : 21 Oct 2020 11:03
 Operator : SY/MD
 Sample : VV1021WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK93

Quant Time: Oct 22 06:20:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 22 06:00:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	303696	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	302861	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	140752	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	113713	45.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.98%
7) Chloroethane-d5	1.58	69	98791	49.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.76%
11) 1,1-Dichloroethene-d2	2.12	63	174014	36.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.74%
21) 2-Butanone-d5	3.91	46	165383	94.91	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.91%
24) Chloroform-d	4.37	84	213741	47.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.18%
26) 1,2-Dichloroethane-d4	5.06	65	147533	49.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.98%
32) Benzene-d6	5.07	84	425825	48.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.32%
36) 1,2-Dichloropropane-d6	6.09	67	142183	50.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.60%
41) Toluene-d8	7.34	98	357658	45.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.18%
43) trans-1,3-Dichloropropene-	7.64	79	64325	46.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.10%
47) 2-Hexanone-d5	8.11	63	87754	83.56	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.56%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	144312	43.56	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.12%
64) 1,2-Dichlorobenzene-d4	11.65	152	147435	51.49	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102120\
Data File : VV019041.D
Acq On : 21 Oct 2020 11:03
Operator : SY/MD
Sample : VV1021WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK93

Quant Time: Oct 22 06:20:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 22 06:00:52 2020
Response via : Initial Calibration

