

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102125\
 Data File : VV039292.D
 Acq On : 21 Oct 2025 09:35
 Operator : SY/MD
 Sample : VV1021MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV1021MBL01

Quant Time: Oct 24 04:37:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.596	168	552746	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	957439	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	890597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	381709	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.937	65	466533	59.399	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	118.800%#
35) Dibromofluoromethane	4.500	113	418389	55.460	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	110.920%
50) Toluene-d8	7.236	98	1226026	47.772	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	95.540%
62) 4-Bromofluorobenzene	10.001	95	412657	45.148	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	90.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102125\
Data File : VV039292.D
Acq On : 21 Oct 2025 09:35
Operator : SY/MD
Sample : VV1021MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV1021MBL01

Quant Time: Oct 24 04:37:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

