

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102125\
 Data File : VV039301.D
 Acq On : 21 Oct 2025 13:37
 Operator : SY/MD
 Sample : Q3390-05DL 500X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 4175DL

Quant Time: Oct 24 04:48:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.597	168	1143315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.529	114	2018397	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1830240	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	843731	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.934	65	832383	51.237	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	102.480%
35) Dibromofluoromethane	4.497	113	728500	45.807	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	91.620%
50) Toluene-d8	7.233	98	2460080	45.470	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	90.940%
62) 4-Bromofluorobenzene	9.998	95	884666	45.913	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	91.820%
Target Compounds						
						Qvalue
16) Acetone	2.118	43	948622	120.323	ug/l	100
52) Toluene	7.310	92	189238	4.890	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102125\
Data File : VV039301.D
Acq On : 21 Oct 2025 13:37
Operator : SY/MD
Sample : Q3390-05DL 500X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
4175DL

Quant Time: Oct 24 04:48:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

