

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102125\
 Data File : VV039302.D
 Acq On : 21 Oct 2025 14:00
 Operator : SY/MD
 Sample : Q3390-06DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 4160DL

Quant Time: Oct 24 04:48:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	723535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1271826	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1215647	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	569982	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.940	65	576159	56.041	ug/l	0.00
Spiked Amount 50.000	Range 81 - 118		Recovery = 112.080%			
35) Dibromofluoromethane	4.503	113	494898	49.386	ug/l	0.00
Spiked Amount 50.000	Range 80 - 119		Recovery = 98.780%			
50) Toluene-d8	7.236	98	1600610	46.951	ug/l	0.00
Spiked Amount 50.000	Range 89 - 112		Recovery = 93.900%			
62) 4-Bromofluorobenzene	10.001	95	597776	49.235	ug/l	0.00
Spiked Amount 50.000	Range 85 - 114		Recovery = 98.480%			
Target Compounds						
30) Chloroform	4.285	83	14697m	0.775	ug/l	
52) Toluene	7.307	92	1671935	68.561	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102125\
Data File : VV039302.D
Acq On : 21 Oct 2025 14:00
Operator : SY/MD
Sample : Q3390-06DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
4160DL

Quant Time: Oct 24 04:48:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

