

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102125\
 Data File : VV039303.D
 Acq On : 21 Oct 2025 15:32
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Oct 24 04:49:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.593	168	689917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.529	114	1212535	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1108899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	491534	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.937	65	512574	52.286	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	104.580%
35) Dibromofluoromethane	4.497	113	451388	47.246	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	94.500%
50) Toluene-d8	7.233	98	1433109	44.093	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	88.180%#
62) 4-Bromofluorobenzene	9.998	95	448257	38.726	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	77.460%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102125\
Data File : VV039303.D
Acq On : 21 Oct 2025 15:32
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Oct 24 04:49:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

