

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102219\
 Data File : VV013274.D
 Acq On : 22 Oct 2019 13:13
 Operator : SY/MD
 Sample : K5321-07
 Misc : 25.00ML/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Oct 23 01:41:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 23 01:23:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	447288	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	495930	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	242308	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	192189	5.83	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	116.60%
7) Chloroethane-d5	1.58	69	158306	5.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.40%
11) 1,1-Dichloroethene-d2	2.13	63	237663	4.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.00%
20) 2-Butanone-d5	3.95	46	321882	50.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.66%
24) Chloroform-d	4.40	84	398044	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
26) 1,2-Dichloroethane-d4	5.08	65	190794	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.20%
32) Benzene-d6	5.10	84	803674	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
36) 1,2-Dichloropropane-d6	6.11	67	240797	4.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.80%
41) Toluene-d8	7.36	98	639285	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
43) trans-1,3-Dichloropropene-	7.66	79	79464	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	8.13	63	240644	36.93	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.86%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	167364	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	256928	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102219\
Data File : VV013274.D
Acq On : 22 Oct 2019 13:13
Operator : SY/MD
Sample : K5321-07
Misc : 25.00ML/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Oct 23 01:41:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 23 01:23:09 2019
Response via : Initial Calibration

