

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102219\
 Data File : VV013279.D
 Acq On : 22 Oct 2019 15:11
 Operator : SY/MD
 Sample : K5462-07MSD
 Misc : 25.00ML/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0G42MSD

Quant Time: Oct 23 02:09:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 23 01:23:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	439857	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	468035	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	246436	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	186328	5.75	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.00%
7) Chloroethane-d5	1.58	69	152630	5.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.00%
11) 1,1-Dichloroethene-d2	2.13	63	281072	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.40%
20) 2-Butanone-d5	3.97	46	470038	75.48	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	150.96%#
24) Chloroform-d	4.40	84	393027	5.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.00%
26) 1,2-Dichloroethane-d4	5.08	65	188037	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.40%
32) Benzene-d6	5.10	84	805415	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.20%
36) 1,2-Dichloropropane-d6	6.12	67	241121	5.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.80%
41) Toluene-d8	7.36	98	662467	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloropropene-	7.66	79	80199	4.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.00%
46) 2-Hexanone-d5	8.13	63	277276	45.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.18%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	169240	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	267977	5.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.80%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.32	62	146187	3.444	ug/L	98
12) 1,1-Dichloroethene	2.14	96	122184	3.533	ug/L	89
17) Methyl tert-butyl Ether	2.81	73	233208	2.912	ug/L	95
18) trans-1,2-Dichloroethene	2.79	96	79545	1.951	ug/L	95
19) 1,1-Dichloroethane	3.23	63	33551	0.506	ug/L	97
22) cis-1,2-Dichloroethene	3.96	96	2429853	55.685	ug/L #	89
33) Benzene	5.14	78	740218	4.142	ug/L	100
34) Trichloroethene	5.96	95	723958	15.096	ug/L	98
42) Toluene	7.43	91	738516	3.935	ug/L	100
47) Tetrachloroethene	8.02	164	14220	0.327	ug/L	96
51) Chlorobenzene	8.92	112	530425	4.317	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102219\
Data File : VV013279.D
Acq On : 22 Oct 2019 15:11
Operator : SY/MD
Sample : K5462-07MSD
Misc : 25.00ML/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0G42MSD

Quant Time: Oct 23 02:09:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 23 01:23:09 2019
Response via : Initial Calibration

