




Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102220\
Data File : VV019078.D
Acq On : 22 Oct 2020 19:23
Operator : SY/MD
Sample : L4471-09
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Quant Time: Oct 23 06:33:38 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 22 13:54:49 2020
Response via : Initial Calibration



TIC: VV019078.D

(13) Acetone (T)

2.206min (+0.006) 9.53ug/L m

response 11563

Ion	Exp%	Act%
43.00	100	100
58.00	30.10	35.20
0.00	0.00	0.00
0.00	0.00	0.00