

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102220\
 Data File : VV019072.D
 Acq On : 22 Oct 2020 17:01
 Operator : SY/MD
 Sample : L4471-13
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG419

Manual Integrations
 APPROVED

MMDadoda
 10/26/2020 1:36:29 PM

Quant Time: Oct 23 07:07:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 22 13:54:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	283874	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	279300	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	134288	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	105830	41.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.08%
7) Chloroethane-d5	1.58	69	84568	40.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.48%
11) 1,1-Dichloroethene-d2	2.12	63	157381	32.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.16%
21) 2-Butanone-d5	3.91	46	130102	75.01	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.01%
24) Chloroform-d	4.37	84	175131	38.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.66%
26) 1,2-Dichloroethane-d4	5.05	65	124013	40.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.80%
32) Benzene-d6	5.07	84	347308	40.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.30%
36) 1,2-Dichloropropane-d6	6.09	67	112522	40.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	81.32%
41) Toluene-d8	7.33	98	307673	40.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.54%
43) trans-1,3-Dichloropropene-	7.64	79	56640	39.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.88%
47) 2-Hexanone-d5	8.10	63	88528	73.57	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.57%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	138320	37.55	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	75.10%
64) 1,2-Dichlorobenzene-d4	11.64	152	119393	42.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.78%

Target Compounds					Ovalue
13) Acetone	2.20	43	4732m	4.025	ug/L
14) Carbon disulfide	2.31	76	33532	5.364	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102220\
Data File : VV019072.D
Acq On : 22 Oct 2020 17:01
Operator : SY/MD
Sample : L4471-13
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG419

Manual Integrations
APPROVED

MMDadoda
10/26/2020 1:36:29 PM

Quant Time: Oct 23 07:07:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 22 13:54:49 2020
Response via : Initial Calibration

