

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102220\
 Data File : VV019080.D
 Acq On : 22 Oct 2020 20:10
 Operator : SY/MD
 Sample : VIBLK66
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK66

Quant Time: Oct 23 07:45:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 22 13:54:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	309010	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	300428	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	139440	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	128372	45.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.46%
7) Chloroethane-d5	1.58	69	106157	46.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.96%
11) 1,1-Dichloroethene-d2	2.12	63	193600	36.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.64%
21) 2-Butanone-d5	3.92	46	163008	86.34	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.34%
24) Chloroform-d	4.38	84	217891	44.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.76%
26) 1,2-Dichloroethane-d4	5.06	65	155761	47.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.38%
32) Benzene-d6	5.07	84	433564	47.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.36%
36) 1,2-Dichloropropane-d6	6.09	67	138582	46.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.10%
41) Toluene-d8	7.34	98	382521	46.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.10%
43) trans-1,3-Dichloropropene-	7.64	79	68384	44.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.66%
47) 2-Hexanone-d5	8.11	63	117818	91.03	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.03%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	178220	44.98	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.96%
64) 1,2-Dichlorobenzene-d4	11.65	152	147647	50.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.96%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102220\
Data File : VV019080.D
Acq On : 22 Oct 2020 20:10
Operator : SY/MD
Sample : VIBLK66
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK66

Quant Time: Oct 23 07:45:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 22 13:54:49 2020
Response via : Initial Calibration

