

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
 Data File : VV022963.D
 Acq On : 21 Oct 2021 18:50
 Operator : SY/MD
 Sample : M4246-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 22 04:56:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 22 04:55:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	146303	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	136643	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	59065	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.298	65	58188	5.781	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.600%
7) Chloroethane-d5	1.558	69	34694	3.836	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.800%
11) 1,1-Dichloroethene-d2	2.098	63	61617	3.121	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.400%
20) 2-Butanone-d5	3.966	46	106384	40.308	ug/L	0.08
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.620%
24) Chloroform-d	4.346	84	91247	4.399	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
26) 1,2-Dichloroethane-d4	5.034	65	43935	4.364	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
32) Benzene-d6	5.043	84	197491	4.855	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	6.072	67	58331	4.842	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.800%
41) Toluene-d8	7.317	98	158558	4.509	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
43) trans-1,3-Dichloroprop...	7.625	79	18032	4.512	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.200%
46) 2-Hexanone-d5	8.104	63	65723	40.811	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.620%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	34011	4.168	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	54593	4.985	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
Data File : VV022963.D
Acq On : 21 Oct 2021 18:50
Operator : SY/MD
Sample : M4246-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 22 04:56:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 22 04:55:17 2021
Response via : Initial Calibration

