

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
 Data File : VV022986.D
 Acq On : 22 Oct 2021 14:52
 Operator : SY/MD
 Sample : M4137-20
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 23 01:28:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 23 01:14:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	132443	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.860	117	128737	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	53282	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	57843	4.998	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 100.000%		
7) Chloroethane-d5	1.564	69	39128	5.466	ug/L	-0.01
Spiked Amount	5.000	Range 65 - 130	Recovery	= 109.400%		
11) 1,1-Dichloroethene-d2	2.114	63	64489	3.863	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 77.200%		
20) 2-Butanone-d5	3.973	46	104739	56.418	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 112.840%		
24) Chloroform-d	4.349	84	91467	4.862	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 97.200%		
26) 1,2-Dichloroethane-d4	5.043	65	46946	5.293	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 105.800%		
32) Benzene-d6	5.047	84	199743	5.314	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 106.200%		
36) 1,2-Dichloropropane-d6	6.098	67	60476	5.227	ug/L	0.01
Spiked Amount	5.000	Range 60 - 140	Recovery	= 104.600%		
41) Toluene-d8	7.329	98	157420	4.663	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 93.200%		
43) trans-1,3-Dichloroprop...	7.638	79	19530	4.818	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 96.400%		
46) 2-Hexanone-d5	8.108	63	67967	45.285	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 90.560%		
56) 1,1,2,2-Tetrachloroeth...	10.220	84	36993	4.630	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 92.600%		
66) 1,2-Dichlorobenzene-d4	11.625	152	54428	5.725	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 114.600%		
Target Compounds						Qvalue
16) Methylene chloride	2.513	84	6247	0.839	ug/L	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
Data File : VV022986.D
Acq On : 22 Oct 2021 14:52
Operator : SY/MD
Sample : M4137-20
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 23 01:28:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 23 01:14:46 2021
Response via : Initial Calibration

