

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
 Data File : VV022997.D
 Acq On : 22 Oct 2021 19:17
 Operator : SY/MD
 Sample : M4265-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB7J1

Quant Time: Oct 23 01:31:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 23 01:14:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	132236	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	130974	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	52178	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	54496	4.716	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	94.400%	
7) Chloroethane-d5	1.561	69	37786	5.287	ug/L	-0.02
Spiked Amount	5.000	Range 65 - 130	Recovery	=	105.800%	
11) 1,1-Dichloroethene-d2	2.101	63	67622	4.057	ug/L	-0.02
Spiked Amount	5.000	Range 60 - 125	Recovery	=	81.200%	
20) 2-Butanone-d5	3.940	46	71759	38.713	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	77.420%	
24) Chloroform-d	4.346	84	78954	4.203	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.000%	
26) 1,2-Dichloroethane-d4	5.037	65	44422	5.017	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.400%	
32) Benzene-d6	5.043	84	194066	5.075	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.600%	
36) 1,2-Dichloropropane-d6	6.072	67	58770	4.993	ug/L	-0.01
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.800%	
41) Toluene-d8	7.317	98	150536	4.383	ug/L	-0.01
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.600%	
43) trans-1,3-Dichloroprop...	7.625	79	18572	4.503	ug/L	-0.01
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.000%	
46) 2-Hexanone-d5	8.104	63	57420	37.604	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	75.200%	
56) 1,1,2,2-Tetrachloroeth...	10.220	84	33270	4.093	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	81.800%	
66) 1,2-Dichlorobenzene-d4	11.625	152	49594	5.327	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
Data File : VV022997.D
Acq On : 22 Oct 2021 19:17
Operator : SY/MD
Sample : M4265-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB7J1

Quant Time: Oct 23 01:31:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 23 01:14:46 2021
Response via : Initial Calibration

