

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
 Data File : VV022961.D
 Acq On : 21 Oct 2021 18:02
 Operator : SY/MD
 Sample : VV1022WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK229

Quant Time: Oct 22 04:55:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 22 04:55:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	141964	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	133237	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58548	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	57110	5.847	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	117.000%
7) Chloroethane-d5	1.561	69	32280	3.678	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.600%
11) 1,1-Dichloroethene-d2	2.098	63	64143	3.348	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.000%
20) 2-Butanone-d5	3.937	46	97178	37.945	ug/L	0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.880%
24) Chloroform-d	4.342	84	93208	4.631	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	5.031	65	42789	4.380	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
32) Benzene-d6	5.040	84	192041	4.842	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	6.069	67	56432	4.804	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.000%
41) Toluene-d8	7.313	98	160893	4.692	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.800%
43) trans-1,3-Dichloroprop...	7.625	79	19047	4.888	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.800%
46) 2-Hexanone-d5	8.098	63	62209	39.616	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.240%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32300	4.059	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	54802	5.048	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
Data File : VV022961.D
Acq On : 21 Oct 2021 18:02
Operator : SY/MD
Sample : VV1022WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK229

Quant Time: Oct 22 04:55:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 22 04:55:17 2021
Response via : Initial Calibration

