

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
 Data File : VV022964.D
 Acq On : 21 Oct 2021 19:14
 Operator : SY/MD
 Sample : M4265-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB7J1

Quant Time: Oct 22 04:56:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 22 04:55:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	123197	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	120961	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	48648	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	4856	0.573	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	11.400%#
7) Chloroethane-d5	1.561	69	7142	0.938	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	18.800%#
11) 1,1-Dichloroethene-d2	2.102	63	4804	0.289	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	5.800%#
20) 2-Butanone-d5	3.941	46	120319	54.137	ug/L	0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.280%
24) Chloroform-d	4.349	84	35696	2.044	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	40.800%#
26) 1,2-Dichloroethane-d4	5.037	65	18926	2.232	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	44.600%#
32) Benzene-d6	5.047	84	33569	0.932	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	18.600%#
36) 1,2-Dichloropropane-d6	6.072	67	19865	1.863	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	37.200%#
41) Toluene-d8	7.320	98	17723	0.569	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	11.400%#
43) trans-1,3-Dichloroprop...	7.632	79	4996	1.412	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	28.200%#
46) 2-Hexanone-d5	8.101	63	58365	40.940	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.880%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26075	3.609	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	72.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	19593	2.172	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	43.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102221\
Data File : VV022964.D
Acq On : 21 Oct 2021 19:14
Operator : SY/MD
Sample : M4265-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB7J1

Quant Time: Oct 22 04:56:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 22 04:55:17 2021
Response via : Initial Calibration

