

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102319\
 Data File : VV013306.D
 Acq On : 23 Oct 2019 20:12
 Operator : SY/MD
 Sample : VV1023WBL02
 Misc : 5.00ML/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK64

Quant Time: Oct 24 04:25:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 24 04:17:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	1322302	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	1274792	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	587752	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	382997	43.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.38%
7) Chloroethane-d5	1.58	69	317706	46.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.92%
11) 1,1-Dichloroethene-d2	2.13	63	477470	37.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.46%
21) 2-Butanone-d5	3.95	46	438469	74.66	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	74.66%
24) Chloroform-d	4.40	84	821404	43.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.98%
26) 1,2-Dichloroethane-d4	5.08	65	514939	45.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.82%
32) Benzene-d6	5.10	84	1759887	44.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.68%
36) 1,2-Dichloropropane-d6	6.12	67	544956	44.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.68%
41) Toluene-d8	7.36	98	1552988	43.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.78%
43) trans-1,3-Dichloropropene-	7.66	79	230768	43.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.06%
47) 2-Hexanone-d5	8.13	63	282409	76.72	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.72%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	699285	43.53	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.06%
64) 1,2-Dichlorobenzene-d4	11.67	152	656543	48.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.36%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102319\
Data File : VV013306.D
Acq On : 23 Oct 2019 20:12
Operator : SY/MD
Sample : VV1023WBL02
Misc : 5.00ML/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK64

Quant Time: Oct 24 04:25:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102319WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 24 04:17:57 2019
Response via : Initial Calibration

