

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102320\
 Data File : VV019099.D
 Acq On : 23 Oct 2020 18:06
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Oct 24 05:28:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 24 03:31:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	262450	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	258462	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	124165	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100059	41.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.94%
7) Chloroethane-d5	1.58	69	85195	44.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.78%
11) 1,1-Dichloroethene-d2	2.12	63	158947	35.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.20%
21) 2-Butanone-d5	3.92	46	140119	87.38	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.38%
24) Chloroform-d	4.38	84	185604	44.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.02%
26) 1,2-Dichloroethane-d4	5.06	65	137085	48.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.80%
32) Benzene-d6	5.07	84	361838	45.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.54%
36) 1,2-Dichloropropane-d6	6.09	67	115536	45.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.22%
41) Toluene-d8	7.34	98	318872	45.10	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	7.64	79	58697	44.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.46%
47) 2-Hexanone-d5	8.11	63	100011	89.82	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.82%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	147758	43.35	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.70%
64) 1,2-Dichlorobenzene-d4	11.65	152	127453	48.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.88%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102320\
Data File : VV019099.D
Acq On : 23 Oct 2020 18:06
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Oct 24 05:28:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 24 03:31:32 2020
Response via : Initial Calibration

