

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW102323\
 Data File : VW032579.D
 Acq On : 23 Oct 2023 18:42
 Operator : SY/MD
 Sample : 04994-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B35

Quant Time: Oct 24 04:48:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 21 02:44:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	85930	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	99494	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	42254	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	25248	3.613	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	72.200%	
7) Chloroethane-d5	1.532	69	26359	4.646	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.000%	
11) 1,1-Dichloroethene-d2	2.060	65	10444	3.544	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	70.800%	
20) 2-Butanone-d5	3.825	46	83814	47.409	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	94.820%	
24) Chloroform-d	4.252	84	57526	4.666	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.400%	
26) 1,2-Dichloroethane-d4	4.947	65	30076	5.150	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.000%	
32) Benzene-d6	4.963	84	96098	3.283	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	65.600%#	
36) 1,2-Dichloropropane-d6	5.989	67	34846	3.592	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	71.800%	
41) Toluene-d8	7.246	98	79806	2.981	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	59.600%#	
43) trans-1,3-Dichloroprop...	7.561	79	11865	3.452	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	69.000%	
46) 2-Hexanone-d5	8.030	63	78459	48.515	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.040%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	33847	4.989	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.800%	
66) 1,2-Dichlorobenzene-d4	11.564	152	34000	4.336	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	86.800%	
Target Compounds						
					Qvalue	
25) Chloroform	4.275	83	108168	8.119	ug/L	97
38) Bromodichloromethane	6.439	83	26481	2.379	ug/L	95
49) Dibromochloromethane	8.185	129	4702	0.683	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102323\
Data File : VV032579.D
Acq On : 23 Oct 2023 18:42
Operator : SY/MD
Sample : 04994-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B35

Quant Time: Oct 24 04:48:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 21 02:44:24 2023
Response via : Initial Calibration

