

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV102418\
 Data File : VV008389.D
 Acq On : 24 Oct 2018 13:23
 Operator : SY/MD
 Sample : VV1024WBL01
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK73

Quant Time: Oct 25 09:01:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 25 08:48:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	223672	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	199821	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	82177	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	53750	43.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.80%
7) Chloroethane-d5	1.59	69	43791	48.97	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.94%
11) 1,1-Dichloroethene-d2	2.13	63	79637	35.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.20%
21) 2-Butanone-d5	3.93	46	131282	97.54	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.54%
24) Chloroform-d	4.41	84	140869	46.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.24%
26) 1,2-Dichloroethane-d4	5.09	65	96891	47.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.90%
32) Benzene-d6	5.10	84	279449	49.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.16%
36) 1,2-Dichloropropane-d6	6.12	67	96775	48.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.92%
41) Toluene-d8	7.36	98	252549	48.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.70%
43) trans-1,3-Dichloropropene-	7.67	79	47753	49.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.46%
47) 2-Hexanone-d5	8.13	63	96788	93.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.78%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	126557	47.28	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.56%
64) 1,2-Dichlorobenzene-d4	11.68	152	82162	48.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.76%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV102418\
Data File : VV008389.D
Acq On : 24 Oct 2018 13:23
Operator : SY/MD
Sample : VV1024WBL01
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK73

Quant Time: Oct 25 09:01:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 25 08:48:14 2018
Response via : Initial Calibration

