

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102418\
 Data File : VV008397.D
 Acq On : 24 Oct 2018 17:28
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD05075

Quant Time: Oct 25 08:54:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 25 08:48:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.357	0.352	1.4	102	0.00
3 T	Chloromethane	0.382	0.356	6.8	97	0.00
4 S	Vinyl Chloride-d3	0.277	0.222	19.9	85	0.00
5 T	Vinyl chloride	0.354	0.352	0.6	104	0.00
6 T	Bromomethane	0.097	0.096	1.0	106	0.00
7 S	Chloroethane-d5	0.200	0.180	10.0	97	0.00
8 T	Chloroethane	0.186	0.193	-3.8	111	0.00
9 T	Trichlorofluoromethane	0.456	0.449	1.5	102	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.238	0.245	-2.9	105	0.00
11 S	1,1-Dichloroethene-d2	0.500	0.444	11.2	93	0.00
12 T	1,1-Dichloroethene	0.224	0.222	0.9	104	0.00
13 T	Acetone	0.234	0.270	-15.4	128	0.00
14 T	Carbon disulfide	1.010	0.991	1.9	103	0.00
15 T	Methyl Acetate	0.466	0.438	6.0	99	0.00
16 T	Methylene chloride	0.389	0.377	3.1	104	0.00
17 T	trans-1,2-Dichloroethene	0.342	0.338	1.2	104	0.00
18 T	Methyl tert-butyl Ether	1.224	1.209	1.2	103	0.00
19 T	1,1-Dichloroethane	0.705	0.694	1.6	102	0.00
20 T	cis-1,2-Dichloroethene	0.390	0.392	-0.5	104	0.00
21 S	2-Butanone-d5	0.301	0.289	4.0	99	0.00
22 T	2-Butanone	0.337	0.353	-4.7	109	0.00
23 T	Bromochloromethane	0.184	0.182	1.1	104	0.00
24 S	Chloroform-d	0.683	0.632	7.5	97	0.00
25 T	Chloroform	0.689	0.685	0.6	104	0.00
26 S	1,2-Dichloroethane-d4	0.457	0.422	7.7	98	0.00
27 T	1,2-Dichloroethane	0.575	0.553	3.8	101	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
29 T	Cyclohexane	0.728	0.718	1.4	106	0.00
30 T	1,1,1-Trichloroethane	0.648	0.631	2.6	103	0.00
31 T	Carbon tetrachloride	0.550	0.540	1.8	104	0.00
32 S	Benzene-d6	1.425	1.285	9.8	96	0.00
33 T	Benzene	1.634	1.604	1.8	103	0.00
34 T	Trichloroethene	0.418	0.408	2.4	104	0.00
35 T	Methylcyclohexane	0.711	0.725	-2.0	106	0.00
36 S	1,2-Dichloropropane-d6	0.495	0.448	9.5	98	0.00
37 T	1,2-Dichloropropane	0.451	0.445	1.3	105	0.00
38 T	Bromodichloromethane	0.575	0.560	2.6	104	0.00
39 T	cis-1,3-Dichloropropene	0.710	0.709	0.1	104	0.00
40 T	4-Methyl-2-pentanone	0.668	0.647	3.1	103	0.00
41 S	Toluene-d8	1.294	1.168	9.7	96	0.00
42 T	Toluene	1.698	1.668	1.8	103	0.00
43 S	trans-1,3-Dichloropropene-d	0.240	0.222	7.5	97	0.00
44 T	trans-1,3-Dichloropropene	0.664	0.661	0.5	104	0.00
45 T	1,1,2-Trichloroethane	0.399	0.391	2.0	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102418\
 Data File : VV008397.D
 Acq On : 24 Oct 2018 17:28
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD05075

Quant Time: Oct 25 08:54:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 25 08:48:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Tetrachloroethene	0.273	0.282	-3.3	108	0.00
47 S	2-Hexanone-d5	0.258	0.251	2.7	102	0.00
48 T	2-Hexanone	0.516	0.537	-4.1	111	0.00
49 T	Dibromochloromethane	0.429	0.418	2.6	101	0.00
50 T	1,2-Dibromoethane	0.419	0.403	3.8	101	0.00
51 T	Chlorobenzene	1.045	1.035	1.0	104	0.00
52 T	Ethylbenzene	1.902	1.911	-0.5	104	0.00
53 T	m,p-Xylene	0.695	0.698	-0.4	104	0.00
54 T	o-xylene	0.694	0.697	-0.4	106	0.00
55 T	Styrene	1.169	1.191	-1.9	106	0.00
56 T	Isopropylbenzene	1.815	1.853	-2.1	106	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.670	0.650	3.0	104	0.00
58 T	1,1,2,2-Tetrachloroethane	0.699	0.706	-1.0	106	0.00
59	1,2,3-Trichloropropane	0.566	0.561	0.9	105	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
61 T	Bromoform	0.631	0.598	5.2	107	0.00
62 T	1,3-Dichlorobenzene	1.653	1.650	0.2	112	0.00
63 T	1,4-Dichlorobenzene	1.657	1.635	1.3	111	0.00
64 S	1,2-Dichlorobenzene-d4	1.023	0.939	8.2	105	0.00
65 T	1,2-Dichlorobenzene	1.686	1.677	0.5	111	0.00
66 T	1,2-Dibromo-3-chloropropane	0.347	0.324	6.6	105	0.00
67	1,3,5-Trichlorobenzene	1.079	1.119	-3.7	115	0.00
68 T	1,2,4-trichlorobenzene	0.873	0.885	-1.4	110	0.00
69	Naphthalene	2.824	2.604	7.8	101	0.00
70 T	1,2,3-Trichlorobenzene	0.873	0.865	0.9	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0