

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW102423\
 Data File : VW032596.D
 Acq On : 24 Oct 2023 17:41
 Operator : SY/MD
 Sample : 04992-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AH9

Quant Time: Oct 25 05:55:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 25 05:52:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	81048	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	87190	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	29204	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	22595	3.428	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	68.600%	
7) Chloroethane-d5	1.532	69	26211	4.899	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	98.000%	
11) 1,1-Dichloroethene-d2	2.059	65	8787	3.161	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	63.200%	
20) 2-Butanone-d5	3.796	46	92770	55.636	ug/L	-0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	111.280%	
24) Chloroform-d	4.252	84	51918	4.465	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.200%	
26) 1,2-Dichloroethane-d4	4.947	65	28595	5.192	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.800%	
32) Benzene-d6	4.960	84	91188	3.554	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	71.000%	
36) 1,2-Dichloropropane-d6	5.992	67	33266	3.914	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	78.200%	
41) Toluene-d8	7.246	98	71747	3.058	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	61.200%#	
43) trans-1,3-Dichloroprop...	7.558	79	10596	3.518	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	70.400%	
46) 2-Hexanone-d5	8.027	63	73396	51.789	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	103.580%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	31732	5.337	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.800%	
66) 1,2-Dichlorobenzene-d4	11.564	152	25189	4.648	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	93.000%	
Target Compounds						
					Qvalue	
13) Acetone	2.130	43	4287	3.336	ug/L	97
14) Carbon disulfide	2.240	76	2905	0.140	ug/L	97
17) Methyl tert-butyl Ether	2.706	73	8148	0.504	ug/L #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102423\
Data File : VV032596.D
Acq On : 24 Oct 2023 17:41
Operator : SY/MD
Sample : 04992-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AH9

Quant Time: Oct 25 05:55:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 25 05:52:45 2023
Response via : Initial Calibration

