

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102423\
 Data File : VV032606.D
 Acq On : 24 Oct 2023 22:08
 Operator : SY/MD
 Sample : VV1024WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK378

Quant Time: Oct 25 08:15:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 25 08:15:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	95243	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	105398	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	45401	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	19281	2.489	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	49.800%
7) Chloroethane-d5	1.532	69	21457	3.413	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.200%
11) 1,1-Dichloroethene-d2	2.060	65	8001	2.449	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.000%#
20) 2-Butanone-d5	3.828	46	91945	46.923	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.840%
24) Chloroform-d	4.252	84	46434	3.398	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	68.000%#
26) 1,2-Dichloroethane-d4	4.947	65	25642	3.962	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.200%
32) Benzene-d6	4.963	84	83521	2.693	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	53.800%#
36) 1,2-Dichloropropane-d6	5.992	67	29366	2.858	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	57.200%#
41) Toluene-d8	7.246	98	68557	2.417	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	48.400%#
43) trans-1,3-Dichloroprop...	7.561	79	10206	2.803	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	56.000%
46) 2-Hexanone-d5	8.030	63	68048	39.721	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.440%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	29739	4.138	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	29528	3.505	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	70.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102423\
Data File : VV032606.D
Acq On : 24 Oct 2023 22:08
Operator : SY/MD
Sample : VV1024WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK378

Quant Time: Oct 25 08:15:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 25 08:15:09 2023
Response via : Initial Calibration

