

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102518\
 Data File : VV008419.D
 Acq On : 25 Oct 2018 19:27
 Operator : SY/MD
 Sample : J5669-10
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETBY9

Quant Time: Oct 26 00:25:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 25 23:17:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	192720	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	173413	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	70871	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	53471	50.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.22%
7) Chloroethane-d5	1.57	69	39868	51.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.50%
11) 1,1-Dichloroethene-d2	2.13	63	70100	36.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.74%
21) 2-Butanone-d5	3.94	46	103676	89.40	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.40%
24) Chloroform-d	4.41	84	118907	45.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.36%
26) 1,2-Dichloroethane-d4	5.09	65	83640	47.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.06%
32) Benzene-d6	5.10	84	245840	49.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.50%
36) 1,2-Dichloropropane-d6	6.12	67	82249	47.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.90%
41) Toluene-d8	7.36	98	221696	49.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.82%
43) trans-1,3-Dichloropropene-	7.67	79	40200	48.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.48%
47) 2-Hexanone-d5	8.14	63	79179	88.40	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.40%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	106465	45.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.66%
64) 1,2-Dichlorobenzene-d4	11.68	152	70166	48.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102518\
Data File : VV008419.D
Acq On : 25 Oct 2018 19:27
Operator : SY/MD
Sample : J5669-10
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETBY9

Quant Time: Oct 26 00:25:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 25 23:17:12 2018
Response via : Initial Calibration

