

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102519\
 Data File : VV013346.D
 Acq On : 25 Oct 2019 15:20
 Operator : SY/MD
 Sample : K5457-04
 Misc : 25.0ML/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0K09

Quant Time: Oct 26 01:24:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 26 00:57:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	445974	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	470215	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	249793	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	182947	5.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.20%
7) Chloroethane-d5	1.58	69	145342	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.20%
11) 1,1-Dichloroethene-d2	2.13	63	253552	4.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.60%
20) 2-Butanone-d5	3.97	46	412916	65.39	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	130.78%#
24) Chloroform-d	4.40	84	358624	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.08	65	173276	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.10	84	793183	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
36) 1,2-Dichloropropane-d6	6.12	67	233278	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.36	98	694104	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
43) trans-1,3-Dichloropropene-	7.66	79	79604	4.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.80%
46) 2-Hexanone-d5	8.14	63	261931	42.40	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.80%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	146898	4.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	263327	5.42	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.40%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chloromethane	1.25	50	5052	0.118	ug/L	90
5) Vinyl chloride	1.32	62	11756	0.273	ug/L #	66
12) 1,1-Dichloroethene	2.14	96	81390	2.321	ug/L #	76
18) trans-1,2-Dichloroethene	2.79	96	7505	0.182	ug/L	92
19) 1,1-Dichloroethane	3.23	63	17812	0.265	ug/L #	92
22) cis-1,2-Dichloroethene	3.96	96	909751	20.563	ug/L #	93
29) 1,1,1-Trichloroethane	4.66	97	7756	0.112	ug/L	97
31) Carbon tetrachloride	4.87	117	14015	0.225	ug/L	98
34) Trichloroethene	5.96	95	958290	19.890	ug/L	96
47) Tetrachloroethene	8.02	164	298093	6.813	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102519\
Data File : VV013346.D
Acq On : 25 Oct 2019 15:20
Operator : SY/MD
Sample : K5457-04
Misc : 25.0ML/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0K09

Quant Time: Oct 26 01:24:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 26 00:57:59 2019
Response via : Initial Calibration

