

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102522\
 Data File : VV028684.D
 Acq On : 25 Oct 2022 12:44
 Operator : SY/MD
 Sample : VIBLK
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK302

Quant Time: Oct 27 04:38:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 27 04:33:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	85768	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	80496	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	36213	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.294	65	57320	7.293	ug/L	0.00
7) Chloroethane-d5	1.555	69	50775	8.067	ug/L	-0.02
11) 1,1-Dichloroethene-d2	2.095	63	82721	5.955	ug/L	-0.02
20) 2-Butanone-d5	3.941	46	31347	19.563	ug/L	-0.03
24) Chloroform-d	4.336	84	92301	7.580	ug/L	-0.02
26) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/L	
32) Benzene-d6	5.040	84	186411	5.686	ug/L	0.00
36) 1,2-Dichloropropane-d6	6.063	67	64992	6.513	ug/L	-0.01
41) Toluene-d8	7.310	98	153141	6.447	ug/L	0.00
43) trans-1,3-Dichloroprop...	7.622	79	17851	6.809	ug/L	0.00
46) 2-Hexanone-d5	8.091	63	89725	81.824	ug/L	-0.02
56) 1,1,2,2-Tetrachloroeth...	10.210	84	40043	7.904	ug/L	0.00
66) 1,2-Dichlorobenzene-d4	11.615	152	48476	7.803	ug/L	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102522\
Data File : VV028684.D
Acq On : 25 Oct 2022 12:44
Operator : SY/MD
Sample : VIBLK
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK302

Quant Time: Oct 27 04:38:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 27 04:33:35 2022
Response via : Initial Calibration

