

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102618\
 Data File : VV008427.D
 Acq On : 26 Oct 2018 12:53
 Operator : SY/MD
 Sample : J5669-11
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETBZ7

Manual Integrations
 APPROVED

MMDadoda
 10/29/2018 11:50:02 AM

Quant Time: Oct 27 05:08:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 27 04:42:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	179135	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	160732	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	65536	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	53730	54.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	108.34%
7) Chloroethane-d5	1.57	69	41679m	58.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	116.40%
11) 1,1-Dichloroethene-d2	2.13	63	73321	40.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.86%
21) 2-Butanone-d5	3.93	46	108957	101.08	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.08%
24) Chloroform-d	4.41	84	124954	51.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.16%
26) 1,2-Dichloroethane-d4	5.09	65	87347	53.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.82%
32) Benzene-d6	5.10	84	259064	56.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.12%
36) 1,2-Dichloropropane-d6	6.12	67	88629	55.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	111.48%
41) Toluene-d8	7.36	98	235536	56.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.26%
43) trans-1,3-Dichloropropene-	7.67	79	41745	54.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.08%
47) 2-Hexanone-d5	8.14	63	83817	100.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.96%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	113393	52.66	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.32%
64) 1,2-Dichlorobenzene-d4	11.68	152	76155	56.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.62%

Target Compounds					Ovalue
13) Acetone	2.19	43	3435	4.106 ug/L	96
20) cis-1,2-Dichloroethene	3.97	96	25905	18.559 ug/L	95
34) Trichloroethene	5.97	95	2928	2.177 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102618\
Data File : VV008427.D
Acq On : 26 Oct 2018 12:53
Operator : SY/MD
Sample : J5669-11
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETBZ7

Manual Integrations
APPROVED

MMDadoda
10/29/2018 11:50:02 AM

Quant Time: Oct 27 05:08:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 27 04:42:32 2018
Response via : Initial Calibration

