

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102618\
 Data File : VV008423.D
 Acq On : 26 Oct 2018 11:07
 Operator : SY/MD
 Sample : VV1026WBL01
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK76

Quant Time: Oct 27 04:51:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 27 04:42:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	229274	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	205863	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	84592	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	64410	50.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.48%
7) Chloroethane-d5	1.58	69	51051	55.70	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.40%
11) 1,1-Dichloroethene-d2	2.13	63	90723	39.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.14%
21) 2-Butanone-d5	3.93	46	134478	97.47	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.47%
24) Chloroform-d	4.40	84	150357	48.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.04%
26) 1,2-Dichloroethane-d4	5.08	65	105652	50.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.94%
32) Benzene-d6	5.10	84	314072	53.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.08%
36) 1,2-Dichloropropane-d6	6.12	67	106877	52.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.96%
41) Toluene-d8	7.36	98	288315	54.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.26%
43) trans-1,3-Dichloropropene-	7.67	79	51705	52.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.52%
47) 2-Hexanone-d5	8.13	63	96434	90.69	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.69%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	134062	48.61	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.22%
64) 1,2-Dichlorobenzene-d4	11.68	152	90735	52.44	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.88%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102618\
Data File : VV008423.D
Acq On : 26 Oct 2018 11:07
Operator : SY/MD
Sample : VV1026WBL01
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK76

Quant Time: Oct 27 04:51:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 27 04:42:32 2018
Response via : Initial Calibration

