

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102618\
 Data File : VV008433.D
 Acq On : 26 Oct 2018 15:13
 Operator : SY/MD
 Sample : J5669-03
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETBX0

Quant Time: Nov 06 07:42:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 27 04:42:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	194619	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	177820	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	71042	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	53040	49.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.44%
7) Chloroethane-d5	1.56	69	41158	52.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.80%
11) 1,1-Dichloroethene-d2	2.13	63	71193	36.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.16%
21) 2-Butanone-d5	3.94	46	114835	98.06	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.06%
24) Chloroform-d	4.41	84	128280	48.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.54%
26) 1,2-Dichloroethane-d4	5.09	65	90001	50.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.30%
32) Benzene-d6	5.10	84	265461	52.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.78%
36) 1,2-Dichloropropane-d6	6.12	67	89991	51.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.32%
41) Toluene-d8	7.36	98	238824	51.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.82%
43) trans-1,3-Dichloropropene-	7.67	79	42440	49.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.32%
47) 2-Hexanone-d5	8.14	63	87345	95.10	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.10%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	116210	48.78	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.56%
64) 1,2-Dichlorobenzene-d4	11.68	152	76530	52.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.32%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102618\
Data File : VV008433.D
Acq On : 26 Oct 2018 15:13
Operator : SY/MD
Sample : J5669-03
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETBX0

Quant Time: Nov 06 07:42:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 27 04:42:32 2018
Response via : Initial Calibration

