

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_V\DATA\VV102619\  
 Data File : VV013375.D  
 Acq On : 26 Oct 2019 15:53  
 Operator : SY/MD  
 Sample : K5567-03  
 Misc : 25.0ML/MSVOA\_V/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 JLNC3

Quant Time: Oct 30 08:01:31 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR102119WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Tue Oct 29 13:11:08 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.66  | 114  | 415935   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.89  | 117  | 464838   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.29 | 152  | 237984   | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |          |
|--------------------------------|--------|-------|----------|----------|------|----------|
| 4) Vinyl Chloride-d3           | 1.32   | 65    | 168868   | 5.51     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 110.20%  |
| 7) Chloroethane-d5             | 1.58   | 69    | 142965   | 5.65     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 113.00%  |
| 11) 1,1-Dichloroethene-d2      | 2.13   | 63    | 210047   | 3.85     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 77.00%   |
| 20) 2-Butanone-d5              | 3.97   | 46    | 363921   | 61.80    | ug/L | 0.02     |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 123.60%  |
| 24) Chloroform-d               | 4.40   | 84    | 378721   | 5.25     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 105.00%  |
| 26) 1,2-Dichloroethane-d4      | 5.08   | 65    | 180520   | 5.40     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 108.00%  |
| 32) Benzene-d6                 | 5.10   | 84    | 791101   | 5.26     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 105.20%  |
| 36) 1,2-Dichloropropane-d6     | 6.12   | 67    | 238325   | 5.27     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 105.40%  |
| 41) Toluene-d8                 | 7.36   | 98    | 687548   | 4.93     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 98.60%   |
| 43) trans-1,3-Dichloropropene- | 7.66   | 79    | 73588    | 4.29     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 85.80%   |
| 46) 2-Hexanone-d5              | 8.14   | 63    | 252659   | 41.37    | ug/L | 0.00     |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 82.74%   |
| 57) 1,1,2,2-Tetrachloroethane- | 10.26  | 84    | 157090   | 4.81     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 96.20%   |
| 64) 1,2-Dichlorobenzene-d4     | 11.67  | 152   | 278966   | 6.03     | ug/L | 0.00     |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 120.60%# |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_V\DATA\VV102619\  
Data File : VV013375.D  
Acq On : 26 Oct 2019 15:53  
Operator : SY/MD  
Sample : K5567-03  
Misc : 25.0ML/MSVOA\_V/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
JLNC3

Quant Time: Oct 30 08:01:31 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR102119WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Tue Oct 29 13:11:08 2019  
Response via : Initial Calibration

