

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV102619\
 Data File : VV013376.D
 Acq On : 26 Oct 2019 16:17
 Operator : SY/MD
 Sample : K5567-04
 Misc : 25.0ML/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JLNC4

Quant Time: Oct 30 08:01:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 29 13:11:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	404664	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	442204	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	222882	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	168174	5.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.80%
7) Chloroethane-d5	1.58	69	135036	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.60%
11) 1,1-Dichloroethene-d2	2.13	63	212149	4.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.00%
20) 2-Butanone-d5	3.94	46	359207	62.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.40%
24) Chloroform-d	4.40	84	347338	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.08	65	170942	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.20%
32) Benzene-d6	5.10	84	743002	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
36) 1,2-Dichloropropane-d6	6.12	67	222106	5.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.20%
41) Toluene-d8	7.36	98	631979	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
43) trans-1,3-Dichloropropene-	7.66	79	67925	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	8.13	63	212384	36.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.12%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	145808	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	247928	5.72	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV102619\
Data File : VV013376.D
Acq On : 26 Oct 2019 16:17
Operator : SY/MD
Sample : K5567-04
Misc : 25.0ML/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JLNC4

Quant Time: Oct 30 08:01:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 29 13:11:08 2019
Response via : Initial Calibration

