

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV102619\
 Data File : VV013380.D
 Acq On : 26 Oct 2019 17:52
 Operator : SY/MD
 Sample : K5567-09
 Misc : 25.0ML/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JLND0

Quant Time: Oct 30 08:03:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 29 13:11:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	464973	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	510332	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	273473	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	162599	4.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.80%
7) Chloroethane-d5	1.58	69	137963	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.60%
11) 1,1-Dichloroethene-d2	2.13	63	201446	3.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.00%
20) 2-Butanone-d5	3.96	46	344162	52.28	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.56%
24) Chloroform-d	4.40	84	352748	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.08	65	169640	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
32) Benzene-d6	5.10	84	754729	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.11	67	225218	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.36	98	667954	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
43) trans-1,3-Dichloropropene-	7.66	79	70121	3.73	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.60%
46) 2-Hexanone-d5	8.14	63	234006	34.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.80%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	146589	4.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	269281	5.06	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.20%

Target Compounds					Qvalue
3) Chloromethane	1.25	50	13870	0.311 ug/L	95
13) Acetone	2.23	43	13871	4.075 ug/L	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV102619\
Data File : VV013380.D
Acq On : 26 Oct 2019 17:52
Operator : SY/MD
Sample : K5567-09
Misc : 25.0ML/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JLND0

Quant Time: Oct 30 08:03:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 29 13:11:08 2019
Response via : Initial Calibration

