

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102620\
 Data File : VV019112.D
 Acq On : 26 Oct 2020 12:15
 Operator : SY/MD
 Sample : L4502-14 100X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG453

Quant Time: Oct 27 07:26:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 27 07:08:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	290282	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	288742	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	135688	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	101597	38.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.06%
7) Chloroethane-d5	1.58	69	83003	39.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.20%
11) 1,1-Dichloroethene-d2	2.12	63	156912	31.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.54%
21) 2-Butanone-d5	3.91	46	146650	82.69	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.69%
24) Chloroform-d	4.38	84	196396	42.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.16%
26) 1,2-Dichloroethane-d4	5.06	65	141101	45.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.00%
32) Benzene-d6	5.07	84	385115	43.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.22%
36) 1,2-Dichloropropane-d6	6.09	67	121652	42.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.04%
41) Toluene-d8	7.34	98	339711	43.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.02%
43) trans-1,3-Dichloropropene-	7.64	79	61236	41.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.54%
47) 2-Hexanone-d5	8.11	63	106430	85.56	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.56%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	158482	41.62	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.24%
64) 1,2-Dichlorobenzene-d4	11.65	152	133917	47.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102620\
Data File : VV019112.D
Acq On : 26 Oct 2020 12:15
Operator : SY/MD
Sample : L4502-14 100X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG453

Quant Time: Oct 27 07:26:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 27 07:08:20 2020
Response via : Initial Calibration

