


Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV102620\
 Data File : VV019107.D
 Acq On : 26 Oct 2020 09:15
 Operator : SY/MD
 Sample : BFB82
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB82

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
 Title : VOC Analysis
 Last Update : Tue Oct 27 07:08:20 2020



AutoFind: Scans 2800, 2801, 2802; Background Corrected with Scan 2789

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	9073	PASS
75	95	30	80	54.9	24461	PASS
95	95	100	100	100.0	44584	PASS
96	95	5	9	6.7	2966	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	84.7	37744	PASS
175	174	5	9	8.2	3103	PASS
176	174	95	101	98.2	37053	PASS
177	176	5	9	7.4	2745	PASS

m/z	Abundance
36.00	562.0
37.00	2669.0
38.00	2212.0
39.10	918.0
40.00	259.0
43.30	110.0
44.00	1201.0
44.90	553.0
47.00	574.0
47.90	368.0
49.00	2068.0
50.10	9196.0
51.00	3206.0
52.00	143.0
54.90	150.0
55.90	719.0
57.00	1343.0
60.00	532.0
61.00	2559.0
62.00	2595.0
63.00	2234.0
64.00	430.0
65.10	564.0
66.90	186.0
68.00	5160.0
69.10	4921.0
70.00	366.0
73.00	2197.0
74.00	8279.0
75.00	24696.0
76.00	2057.0
77.00	311.0
78.00	167.0
78.90	1874.0
80.00	501.0
81.00	1809.0
81.90	371.0
87.00	1674.0
88.00	1810.0
90.80	247.0
92.00	1492.0
93.00	2361.0
94.10	5684.0
95.00	44248.0
96.00	3026.0
103.90	187.0
105.80	220.0
114.90	103.0
115.90	259.0
116.80	433.0
117.90	202.0
118.90	319.0
127.90	221.0
129.80	243.0
140.90	549.0
141.90	102.0
142.90	837.0
145.00	278.0
154.90	124.0
171.10	124.0
172.00	538.0
174.00	36568.0
175.00	3038.0
176.00	35936.0
177.00	2657.0
177.80	116.0

Instrument :
MSVOA_V
ClientSampleId :
BFB82

m/z	Abundance
36.00	569.0
37.00	2679.0
38.00	2224.0
39.10	1005.0
39.90	262.0
44.00	1249.0
45.00	490.0
47.00	558.0
47.90	397.0
49.00	2180.0
50.00	9409.0
51.10	3119.0
52.00	164.0
54.90	170.0
56.00	799.0
57.00	1408.0
60.00	491.0
61.00	2544.0
62.00	2670.0
63.00	2320.0
64.00	391.0
65.10	639.0
66.90	177.0
68.00	5250.0
69.10	5203.0
69.90	344.0
71.90	238.0
73.00	2191.0
74.00	8634.0
75.00	25312.0
76.00	2010.0
77.00	392.0
78.00	203.0
78.90	1844.0
80.00	567.0
80.90	1862.0
81.90	368.0
87.00	1639.0
88.00	1769.0
90.80	287.0
92.00	1497.0
93.00	2416.0
94.10	5837.0
95.00	46000.0
96.00	3078.0
103.80	193.0
106.00	217.0
114.90	110.0
115.90	231.0
116.80	425.0
117.90	213.0
118.90	348.0
127.90	243.0
129.80	192.0
134.80	112.0
140.90	572.0
141.90	111.0
142.90	899.0
145.00	320.0
145.80	104.0
148.00	105.0
155.00	133.0
171.00	121.0
171.90	680.0
174.00	39000.0
175.00	3241.0
176.00	38264.0
177.00	2875.0
177.80	118.0

Instrument :
MSVOA_V
ClientSampleId :
BFB82

m/z	Abundance
36.00	488.0
37.00	2445.0
38.10	2105.0
39.10	1028.0
39.90	239.0
44.00	1200.0
45.00	405.0
47.10	510.0
48.00	379.0
49.00	1959.0
50.00	8614.0
51.10	2738.0
52.00	145.0
54.90	164.0
56.00	785.0
57.00	1282.0
60.00	477.0
61.00	2319.0
62.00	2470.0
63.00	2087.0
64.10	350.0
65.10	641.0
66.90	148.0
68.00	4689.0
69.00	4939.0
69.90	334.0
72.00	273.0
72.30	286.0
73.00	1954.0
74.00	8096.0
75.00	23376.0
76.00	1822.0
77.00	398.0
78.00	216.0
78.90	1759.0
80.00	565.0
80.90	1712.0
81.90	360.0
87.00	1491.0
88.00	1586.0
90.80	256.0
92.00	1414.0
93.00	2179.0
94.10	5369.0
95.00	43504.0
96.00	2796.0
103.80	184.0
106.00	206.0
115.90	188.0
116.90	388.0
117.90	194.0
118.90	312.0
128.00	225.0
129.90	155.0
134.80	133.0
140.90	592.0
142.90	870.0
145.00	326.0
145.80	109.0
148.00	121.0
155.10	140.0
171.00	122.0
171.90	719.0
174.00	37664.0
175.00	3032.0
176.00	36960.0
177.00	2705.0

Instrument :
MSVOA_V
ClientSampleId :
BFB82

Scan 2789 (10.058 min): VV019107.D
BFB82

m/z	Abundance
39.90	189.0
44.00	776.0

Instrument :
MSVOA_V
ClientSampleId :
BFB82