




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102621\
 Data File : VV023040.D
 Acq On : 26 Oct 2021 15:39
 Operator : SY/MD
 Sample : M4277-10DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Quant Time: Oct 27 01:33:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 27 01:29:33 2021
 Response via : Initial Calibration






TIC: VV023040.D\data.ms

(13) Acetone (T)**2.195min (+ 0.003) 24.46 ug/L****response 22419**

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	27.70	16.04
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102621\
Data File : VV023040.D
Acq On : 26 Oct 2021 15:39
Operator : SY/MD
Sample : M4277-10DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Quant Time: Oct 27 01:33:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 27 01:29:33 2021
Response via : Initial Calibration



TIC: VV023040.D\data.ms

(13) Acetone (T)**2.195min (+ 0.003) 39.97 ug/L m****response 36642**

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	27.70	9.81
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102621\
 Data File : VV023040.D
 Acq On : 26 Oct 2021 15:39
 Operator : SY/MD
 Sample : M4277-10DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Quant Time: Oct 27 01:33:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 27 01:29:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	133298	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	132565	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56733	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	49047	4.211	ug/L	0.00
Spiked Amount 5.000	Range 40	- 130	Recovery	=	84.200%	
7) Chloroethane-d5	1.568	69	28615	3.972	ug/L	0.00
Spiked Amount 5.000	Range 65	- 130	Recovery	=	79.400%	
11) 1,1-Dichloroethene-d2	2.108	63	53801	3.202	ug/L	0.00
Spiked Amount 5.000	Range 60	- 125	Recovery	=	64.000%	
20) 2-Butanone-d5	3.912	46	79313	42.448	ug/L	0.00
Spiked Amount 50.000	Range 40	- 130	Recovery	=	84.900%	
24) Chloroform-d	4.352	84	87064	4.598	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125	Recovery	=	92.000%	
26) 1,2-Dichloroethane-d4	5.034	65	45368	5.083	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130	Recovery	=	101.600%	
32) Benzene-d6	5.053	84	192247	4.967	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125	Recovery	=	99.400%	
36) 1,2-Dichloropropane-d6	6.072	67	59620	5.004	ug/L	0.00
Spiked Amount 5.000	Range 60	- 140	Recovery	=	100.000%	
41) Toluene-d8	7.317	98	162062	4.662	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130	Recovery	=	93.200%	
43) trans-1,3-Dichloroprop...	7.625	79	19793	4.742	ug/L	0.00
Spiked Amount 5.000	Range 55	- 130	Recovery	=	94.800%	
46) 2-Hexanone-d5	8.095	63	75178	48.643	ug/L	0.00
Spiked Amount 50.000	Range 45	- 130	Recovery	=	97.280%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39936	4.854	ug/L	0.00
Spiked Amount 5.000	Range 65	- 120	Recovery	=	97.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	59216	5.850	ug/L	0.00
Spiked Amount 5.000	Range 80	- 120	Recovery	=	117.000%	
Target Compounds						
					Qvalue	
13) Acetone	2.195	43	36642m	39.971	ug/L	
16) Methylene chloride	2.507	84	3728	0.497	ug/L	95
17) Methyl tert-butyl Ether	2.767	73	887	0.054	ug/L #	55
22) cis-1,2-Dichloroethene	3.918	96	2449	0.296	ug/L #	84
25) Chloroform	4.381	83	12208	0.690	ug/L	91
34) Trichloroethene	5.928	95	1502	0.168	ug/L	92
47) Tetrachloroethene	7.979	164	7799	1.030	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102621\
Data File : VV023040.D
Acq On : 26 Oct 2021 15:39
Operator : SY/MD
Sample : M4277-10DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Quant Time: Oct 27 01:33:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 27 01:29:33 2021
Response via : Initial Calibration

