

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW102623\
 Data File : VW032653.D
 Acq On : 26 Oct 2023 11:30
 Operator : SY/MD
 Sample : 04961-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 27 05:34:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 27 05:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	87095	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	99779	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	43060	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	31111	4.392	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.800%
7) Chloroethane-d5	1.532	69	29558	5.141	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.800%
11) 1,1-Dichloroethene-d2	2.060	65	12582	4.212	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.200%
20) 2-Butanone-d5	3.809	46	50457	28.159	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	56.320%
24) Chloroform-d	4.253	84	58557	4.686	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
26) 1,2-Dichloroethane-d4	4.947	65	29944	5.059	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	4.963	84	113982	3.882	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.600%
36) 1,2-Dichloropropane-d6	5.992	67	36486	3.751	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.000%
41) Toluene-d8	7.246	98	102828	3.830	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.600%
43) trans-1,3-Dichloroprop...	7.558	79	13563	3.935	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.600%
46) 2-Hexanone-d5	8.030	63	64444	39.735	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.480%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	32685	4.804	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	11.564	152	38749	4.849	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102623\
Data File : VV032653.D
Acq On : 26 Oct 2023 11:30
Operator : SY/MD
Sample : 04961-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 27 05:34:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 27 05:30:38 2023
Response via : Initial Calibration

