

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102623\
 Data File : VV032651.D
 Acq On : 26 Oct 2023 10:41
 Operator : SY/MD
 Sample : VV1026WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK381

Quant Time: Oct 27 05:33:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 27 05:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	92877	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	106450	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	43658	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	32991	4.368	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.400%
7) Chloroethane-d5	1.529	69	31361	5.115	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.200%
11) 1,1-Dichloroethene-d2	2.056	65	13909	4.366	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.400%
20) 2-Butanone-d5	3.828	46	80031	41.883	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.760%
24) Chloroform-d	4.252	84	56973	4.276	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
26) 1,2-Dichloroethane-d4	4.947	65	30387	4.814	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	4.963	84	114907	3.669	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.400%
36) 1,2-Dichloropropane-d6	5.992	67	36171	3.485	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	69.800%
41) Toluene-d8	7.246	98	102992	3.596	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.000%
43) trans-1,3-Dichloroprop...	7.561	79	13972	3.799	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.000%
46) 2-Hexanone-d5	8.034	63	65257	37.715	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.420%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	32295	4.449	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.000%
66) 1,2-Dichlorobenzene-d4	11.564	152	38138	4.708	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102623\
Data File : VV032651.D
Acq On : 26 Oct 2023 10:41
Operator : SY/MD
Sample : VV1026WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK381

Quant Time: Oct 27 05:33:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 27 05:30:38 2023
Response via : Initial Calibration

