

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102720\
 Data File : VV019127.D
 Acq On : 27 Oct 2020 11:39
 Operator : SY/MD
 Sample : VV1027MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK99

Quant Time: Oct 28 01:16:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 28 01:06:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	265838	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	259798	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	124148	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80568	33.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.72%
7) Chloroethane-d5	1.58	69	73892	38.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	76.02%
11) 1,1-Dichloroethene-d2	2.12	63	131886	29.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.32%#
21) 2-Butanone-d5	3.91	46	149171	91.84	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.84%
24) Chloroform-d	4.38	84	177985	42.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.28%
26) 1,2-Dichloroethane-d4	5.06	65	130649	46.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.02%
32) Benzene-d6	5.07	84	332401	41.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.66%
36) 1,2-Dichloropropane-d6	6.09	67	108136	42.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.02%
41) Toluene-d8	7.34	98	289938	40.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.60%
43) trans-1,3-Dichloropropene-	7.64	79	55649	42.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.38%
47) 2-Hexanone-d5	8.11	63	105117	93.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.92%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	149232	43.56	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.12%
64) 1,2-Dichlorobenzene-d4	11.65	152	124729	47.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102720\
Data File : VV019127.D
Acq On : 27 Oct 2020 11:39
Operator : SY/MD
Sample : VV1027MBL01
Misc : 5.0µL/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK99

Quant Time: Oct 28 01:16:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102220WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 28 01:06:58 2020
Response via : Initial Calibration

