

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW102723\
 Data File : VW032701.D
 Acq On : 27 Oct 2023 22:10
 Operator : SY/MD
 Sample : 05055-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A49G3

Quant Time: Oct 28 00:29:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 28 00:17:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	134819	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	156252	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	68206	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	51301	4.633	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.600%
7) Chloroethane-d5	1.532	69	49462	5.257	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.200%
11) 1,1-Dichloroethene-d2	2.060	65	22682	4.564	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.200%
20) 2-Butanone-d5	3.802	46	97395	40.733	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.460%
24) Chloroform-d	4.256	84	94842	5.159	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.200%
26) 1,2-Dichloroethane-d4	4.947	65	48889	5.386	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.800%
32) Benzene-d6	4.963	84	183231	4.925	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
36) 1,2-Dichloropropane-d6	5.992	67	56656	5.259	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%
41) Toluene-d8	7.246	98	169608	4.498	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%
43) trans-1,3-Dichloroprop...	7.561	79	20737	4.362	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.200%
46) 2-Hexanone-d5	8.030	63	79033	41.953	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.900%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	45016	4.699	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	11.564	152	61211	5.496	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.000%
Target Compounds						
					Qvalue	
3) Chloromethane	1.217	50	10577	0.648	ug/L	99
13) Acetone	2.130	43	4800	2.034	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102723\
Data File : VV032701.D
Acq On : 27 Oct 2023 22:10
Operator : SY/MD
Sample : 05055-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A49G3

Quant Time: Oct 28 00:29:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 28 00:17:24 2023
Response via : Initial Calibration

