

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102723\
 Data File : VV032731.D
 Acq On : 28 Oct 2023 10:54
 Operator : SY/MD
 Sample : VV1027WBL03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK384

Quant Time: Oct 30 01:51:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 30 01:47:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	183073	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	193604	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	80144	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	46803	3.113	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.200%
7) Chloroethane-d5	1.532	69	49477	3.872	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.400%
11) 1,1-Dichloroethene-d2	2.056	65	21429	3.175	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.600%
20) 2-Butanone-d5	3.805	46	144771	44.588	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.180%
24) Chloroform-d	4.246	84	104582	4.189	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
26) 1,2-Dichloroethane-d4	4.944	65	54770	4.443	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.800%
32) Benzene-d6	4.960	84	194155	4.212	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	5.989	67	64110	4.803	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.000%
41) Toluene-d8	7.243	98	165417	3.540	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.800%
43) trans-1,3-Dichloroprop...	7.558	79	19954	3.388	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.800%
46) 2-Hexanone-d5	8.027	63	116176	49.771	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.540%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	54814	4.618	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.400%
66) 1,2-Dichlorobenzene-d4	11.564	152	64017	4.892	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102723\
Data File : VV032731.D
Acq On : 28 Oct 2023 10:54
Operator : SY/MD
Sample : VV1027WBL03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 63 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK384

Quant Time: Oct 30 01:51:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 30 01:47:59 2023
Response via : Initial Calibration

