

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102821\
 Data File : VV023082.D
 Acq On : 28 Oct 2021 18:30
 Operator : SY/MD
 Sample : VV1028MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK264

Quant Time: Oct 29 02:15:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 29 01:52:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	291776	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	281098	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	135780	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	93039	46.269	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	92.540%		
7) Chloroethane-d5	1.568	69	72764	46.394	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	92.780%		
11) 1,1-Dichloroethene-d2	2.108	63	132900	34.891	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	69.780%		
21) 2-Butanone-d5	3.879	46	157578	104.596	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	104.600%		
24) Chloroform-d	4.349	84	194658	46.042	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.080%		
26) 1,2-Dichloroethane-d4	5.034	65	123610	49.396	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.800%		
32) Benzene-d6	5.053	84	385646	50.082	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.160%		
36) 1,2-Dichloropropane-d6	6.069	67	122517	49.981	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	99.960%		
41) Toluene-d8	7.317	98	338817	48.228	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.460%		
43) trans-1,3-Dichloroprop...	7.622	79	55440	47.187	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	94.380%		
47) 2-Hexanone-d5	8.088	63	117717	110.566	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	110.570%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	174097	48.593	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	97.180%		
66) 1,2-Dichlorobenzene-d4	11.625	152	139581	53.293	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	106.580%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102821\
Data File : VV023082.D
Acq On : 28 Oct 2021 18:30
Operator : SY/MD
Sample : VV1028MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK264

Quant Time: Oct 29 02:15:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 29 01:52:27 2021
Response via : Initial Calibration

