

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\

Method File : SFAMVLM102821WMA.M

Title : VOC Analysis

Last Update : Sat Oct 30 00:25:16 2021

Response Via : Initial Calibration

Calibration Files

5 =VV023076.D 10 =VV023077.D 50 =VV023092.D 100 =VV023079.D 200 =VV023080.D

	Compound	5	10	50	100	200	Avg	%RSD
<hr/>								
1) I	1,4-Difluorobenzene			-----ISTD-----				
2) T	Dichlorodifluoro...	0.497	0.503	0.485	0.442	0.450	0.475	5.80
3) T	Chloromethane	0.473	0.492	0.445	0.414	0.417	0.448	7.58
4) S	Vinyl Chloride-d3	0.356	0.359	0.336	0.331	0.340	0.345	3.56
5) T	Vinyl chloride	0.455	0.475	0.453	0.424	0.433	0.448	4.40
6) T	Bromomethane	0.277	0.287	0.249	0.223	0.222	0.252	12.01
7) S	Chloroethane-d5	0.288	0.301	0.260	0.261	0.234	0.269	9.89
8) T	Chloroethane	0.298	0.299	0.271	0.252	0.244	0.273	9.26
9) T	Trichlorofluorom...	0.715	0.731	0.683	0.629	0.641	0.680	6.56
10) T	1,1,2-Trichloro....	0.385	0.382	0.366	0.335	0.337	0.361	6.60
11) S	1,1-Dichloroethe...	0.681	0.684	0.639	0.623	0.636	0.653	4.25
12) T	1,1-Dichloroethene	0.347	0.365	0.331	0.312	0.320	0.335	6.34
13) T	Acetone	0.283	0.267	0.266	0.235	0.230	0.256	8.83
14) T	Carbon disulfide	0.990	1.056	1.010	0.961	0.991	1.002	3.50
15) T	Methyl Acetate	0.430	0.439	0.425	0.413	0.417	0.425	2.47
16) T	Methylene chloride	0.456	0.462	0.426	0.399	0.402	0.429	6.83
17) T	trans-1,2-Dichlo...	0.376	0.399	0.390	0.369	0.377	0.382	3.20
18) T	Methyl tert-butyl...	1.052	1.136	1.184	1.158	1.209	1.148	5.27
19) T	1,1-Dichloroethane	0.709	0.772	0.717	0.675	0.687	0.712	5.27
20) T	cis-1,2-Dichloro...	0.372	0.411	0.423	0.411	0.425	0.409	5.25
21) S	2-Butanone-d5	0.231	0.243	0.258	0.277	0.281	0.258	8.34
22) T	2-Butanone	0.254	0.267	0.314	0.311	0.311	0.292	9.75
23) T	Bromochloromethane	0.229	0.243	0.232	0.222	0.225	0.230	3.40
24) S	Chloroform-d	0.709	0.768	0.720	0.709	0.716	0.724	3.40
25) T	Chloroform	0.729	0.775	0.748	0.698	0.711	0.732	4.17
26) S	1,2-Dichloroetha...	0.413	0.437	0.428	0.428	0.438	0.429	2.39
27) T	1,2-Dichloroethane	0.550	0.573	0.558	0.528	0.540	0.550	3.13
28) I	Chlorobenzene-d5			-----ISTD-----				
29) T	Cyclohexane	0.502	0.550	0.622	0.620	0.625	0.584	9.50
30) T	1,1,1-Trichloroe...	0.661	0.701	0.671	0.649	0.650	0.666	3.20
31) T	Carbon tetrachloro...	0.594	0.628	0.606	0.583	0.583	0.599	3.15
32) S	Benzene-d6	1.281	1.377	1.396	1.400	1.395	1.370	3.69
33) T	Benzene	1.463	1.569	1.635	1.562	1.559	1.558	3.93
34) T	Trichloroethene	0.406	0.411	0.408	0.402	0.408	0.407	0.81
35) T	Methylcyclohexane	0.561	0.605	0.677	0.661	0.669	0.634	7.87
36) S	1,2-Dichloroprop...	0.417	0.451	0.432	0.443	0.437	0.436	2.97
37) T	1,2-Dichloropropane	0.403	0.416	0.422	0.404	0.402	0.409	2.18
38) T	Bromodichloromet...	0.539	0.575	0.557	0.547	0.551	0.554	2.47
39) T	cis-1,3-Dichloro...	0.525	0.593	0.655	0.656	0.675	0.621	10.00
40) T	4-Methyl-2-penta...	0.439	0.512	0.562	0.572	0.555	0.528	10.43
41) S	Toluene-d8	1.105	1.245	1.294	1.307	1.298	1.250	6.76
42) T	Toluene	1.467	1.691	1.781	1.703	1.693	1.667	7.08
43) S	trans-1,3-Dichlo...	0.183	0.197	0.215	0.222	0.228	0.209	8.88
44) T	trans-1,3-Dichlo...	0.510	0.565	0.639	0.639	0.656	0.602	10.33
45) T	1,1,2-Trichloroe...	0.413	0.447	0.423	0.406	0.400	0.418	4.39
46) T	Tetrachloroethene	0.343	0.370	0.360	0.340	0.342	0.351	3.85
47) S	2-Hexanone-d5	0.137	0.171	0.199	0.222	0.219	0.189	18.96
48) T	2-Hexanone	0.351	0.388	0.459	0.459	0.439	0.419	11.44
49) T	Dibromochloromet...	0.464	0.484	0.486	0.478	0.477	0.478	1.85
50) T	1,2-Dibromoethane	0.420	0.459	0.447	0.436	0.439	0.440	3.20
51) T	Chlorobenzene	1.093	1.180	1.146	1.104	1.108	1.126	3.21
52) T	Ethylbenzene	1.542	1.750	1.876	1.844	1.857	1.774	7.81
53) T	m,p-Xylene	0.580	0.651	0.743	0.719	0.725	0.684	9.86
54) T	o-Xylene	0.564	0.637	0.723	0.702	0.702	0.665	9.78
55) T	Styrene	0.930	1.094	1.286	1.247	1.227	1.157	12.59
56) S	1,1,2,2-Tetrachl...	0.636	0.658	0.632	0.643	0.618	0.637	2.30

Response Factor Report MSVOA_V

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\

Method File : SFAMVLM102821WMA.M

57)	T	1,1,2,2-Tetrachloroethane	0.656 0.715 0.677 0.650 0.635 0.667	4.61
58)	I	1,4-Dichlorobenzene	-----ISTD-----	
59)	T	Bromoform	0.611 0.641 0.636 0.629 0.646 0.632	2.11
60)		Isopropylbenzene	2.798 3.127 3.401 3.286 3.360 3.194	7.67
61)		1,2,3-Trichloropropane	0.950 0.997 0.941 0.901 0.895 0.937	4.41
62)		1,3,5-Trimethylbenzene	2.147 2.423 2.850 2.825 2.908 2.631	12.61
63)		1,2,4-Trimethylbenzene	2.095 2.394 2.907 2.876 2.934 2.641	14.31
64)	T	1,3-Dichlorobenzene	1.527 1.635 1.681 1.600 1.630 1.615	3.51
65)	T	1,4-Dichlorobenzene	1.673 1.754 1.708 1.633 1.641 1.682	2.97
66)	S	1,2-Dichlorobenzene	0.985 0.974 0.941 0.951 0.970 0.964	1.87
67)	T	1,2-Dichlorobenzene	1.594 1.689 1.664 1.602 1.615 1.633	2.55
68)	T	1,2-Dibromo-3-chloropropane	0.259 0.255 0.255 0.265 0.274 0.262	3.04
69)		1,3,5-Trichlorobutane	1.144 1.202 1.246 1.225 1.290 1.221	4.42
70)	T	1,2,4-trichlorobutane	0.901 0.989 1.095 1.111 1.196 1.059	10.84
71)		Naphthalene	2.639 2.934 3.653 3.846 3.921 3.399	16.97
72)	T	1,2,3-Trichlorobutane	0.978 1.089 1.164 1.166 1.198 1.119	7.89

(#) = Out of Range