

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV102918\
 Data File : VV008446.D
 Acq On : 29 Oct 2018 11:57
 Operator : SY/MD
 Sample : VV1029WBL01
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK77

Quant Time: Oct 30 07:50:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 30 07:40:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	206311	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	184030	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	73550	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	58273	51.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.02%
7) Chloroethane-d5	1.59	69	47613	57.73	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	115.46%
11) 1,1-Dichloroethene-d2	2.13	63	84075	40.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.50%
21) 2-Butanone-d5	3.93	46	131766	106.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.14%
24) Chloroform-d	4.41	84	149055	52.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.82%
26) 1,2-Dichloroethane-d4	5.08	65	101724	54.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.00%
32) Benzene-d6	5.10	84	300249	57.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.50%
36) 1,2-Dichloropropane-d6	6.12	67	103802	57.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	114.04%
41) Toluene-d8	7.36	98	273794	57.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.00%
43) trans-1,3-Dichloropropene-	7.67	79	49340	55.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	111.58%
47) 2-Hexanone-d5	8.13	63	93634	98.51	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.51%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	131633	53.39	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.78%
64) 1,2-Dichlorobenzene-d4	11.68	152	86842	57.72	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.44%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV102918\
Data File : VV008446.D
Acq On : 29 Oct 2018 11:57
Operator : SY/MD
Sample : VV1029WBL01
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK77

Quant Time: Oct 30 07:50:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 30 07:40:49 2018
Response via : Initial Calibration

