

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102920\
 Data File : VV019170.D
 Acq On : 30 Oct 2020 01:33
 Operator : SY/MD
 Sample : L4502-20ME 4X
 Misc : 6.07g/5.0ml/100uL/5.0ml/MSVOA-V/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG443ME

Quant Time: Oct 30 06:23:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 30 05:31:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	239741	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	230609	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	121823	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	81980	43.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.06%
7) Chloroethane-d5	1.58	69	67731	45.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.52%
11) 1,1-Dichloroethene-d2	2.12	63	126222	34.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.66%
21) 2-Butanone-d5	3.91	46	110652	94.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.83%
24) Chloroform-d	4.37	84	161005	45.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.86%
26) 1,2-Dichloroethane-d4	5.05	65	117712	47.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.34%
32) Benzene-d6	5.07	84	312411	47.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.52%
36) 1,2-Dichloropropane-d6	6.09	67	94095	48.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.48%
41) Toluene-d8	7.34	98	291797	46.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.86%
43) trans-1,3-Dichloropropene-	7.64	79	50072	44.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.72%
47) 2-Hexanone-d5	8.11	63	82284	97.41	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.41%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	123454	46.94	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.88%
64) 1,2-Dichlorobenzene-d4	11.65	152	127834	49.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.88%

Target Compounds					Ovalue
46) Tetrachloroethene	8.00	164	50810	32.668 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102920\
Data File : VV019170.D
Acq On : 30 Oct 2020 01:33
Operator : SY/MD
Sample : L4502-20ME 4X
Misc : 6.07g/5.0ml/100uL/5.0ml/MSVOA-V/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG443ME

Quant Time: Oct 30 06:23:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 30 05:31:22 2020
Response via : Initial Calibration

