

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102921\
 Data File : VV023094.D
 Acq On : 29 Oct 2021 10:48
 Operator : SY/MD
 Sample : VV1029MBL04
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK269

Quant Time: Oct 30 00:25:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 30 00:25:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	263181	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	255745	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	124080	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	82579	45.529	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.060%
7) Chloroethane-d5	1.564	69	67206	47.506	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.020%
11) 1,1-Dichloroethene-d2	2.105	63	119746	34.853	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.700%
21) 2-Butanone-d5	3.876	46	132374	97.413	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.410%
24) Chloroform-d	4.349	84	179506	47.071	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.140%
26) 1,2-Dichloroethane-d4	5.030	65	116093	51.433	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.860%
32) Benzene-d6	5.050	84	349039	49.822	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.640%
36) 1,2-Dichloropropane-d6	6.069	67	112054	50.244	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.480%
41) Toluene-d8	7.317	98	305667	47.822	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.640%
43) trans-1,3-Dichloroprop...	7.622	79	52062	48.705	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.400%
47) 2-Hexanone-d5	8.088	63	93744	96.778	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.780%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	157521	48.325	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.660%
66) 1,2-Dichlorobenzene-d4	11.625	152	127389	53.224	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102921\
Data File : VV023094.D
Acq On : 29 Oct 2021 10:48
Operator : SY/MD
Sample : VV1029MBL04
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK269

Quant Time: Oct 30 00:25:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 30 00:25:16 2021
Response via : Initial Calibration

