

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102921\
 Data File : VV023096.D
 Acq On : 29 Oct 2021 11:34
 Operator : SY/MD
 Sample : VV1029WBL03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK271

Quant Time: Oct 30 00:26:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 30 00:25:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	268411	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	266862	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	126725	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	85509	46.226	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.460%
7) Chloroethane-d5	1.568	69	69477	48.154	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.300%
11) 1,1-Dichloroethene-d2	2.108	63	121004	34.533	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.060%
21) 2-Butanone-d5	3.879	46	138676	100.062	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.060%
24) Chloroform-d	4.352	84	186100	47.850	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.700%
26) 1,2-Dichloroethane-d4	5.034	65	117354	50.979	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.960%
32) Benzene-d6	5.053	84	357150	48.856	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.720%
36) 1,2-Dichloropropane-d6	6.072	67	113493	48.770	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.540%
41) Toluene-d8	7.320	98	316126	47.398	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.800%
43) trans-1,3-Dichloroprop...	7.622	79	52785	47.324	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.640%
47) 2-Hexanone-d5	8.088	63	99433	98.375	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.370%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	162920	47.899	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	132885	54.362	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102921\
Data File : VV023096.D
Acq On : 29 Oct 2021 11:34
Operator : SY/MD
Sample : VV1029WBL03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK271

Quant Time: Oct 30 00:26:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 30 00:25:16 2021
Response via : Initial Calibration

