

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV103018\
 Data File : VV008468.D
 Acq On : 30 Oct 2018 15:12
 Operator : SY/MD
 Sample : VV1030WBL01
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK78

Quant Time: Oct 31 07:47:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 31 07:14:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	202139	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	178149	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	72432	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	43420	38.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.60%
7) Chloroethane-d5	1.58	69	38188	47.26	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.52%
11) 1,1-Dichloroethene-d2	2.13	63	65040	32.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.34%
21) 2-Butanone-d5	3.93	46	118707	97.59	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.59%
24) Chloroform-d	4.41	84	124252	45.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.02%
26) 1,2-Dichloroethane-d4	5.08	65	86907	47.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.18%
32) Benzene-d6	5.10	84	247230	48.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.40%
36) 1,2-Dichloropropane-d6	6.12	67	87067	49.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.82%
41) Toluene-d8	7.36	98	219256	47.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.14%
43) trans-1,3-Dichloropropene-	7.67	79	40925	47.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.60%
47) 2-Hexanone-d5	8.13	63	85429	92.84	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.84%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	113386	47.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.02%
64) 1,2-Dichlorobenzene-d4	11.68	152	72057	48.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV103018\
Data File : VV008468.D
Acq On : 30 Oct 2018 15:12
Operator : SY/MD
Sample : VV1030WBL01
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK78

Quant Time: Oct 31 07:47:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102318WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 31 07:14:32 2018
Response via : Initial Calibration

