

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV103020\
 Data File : VV019182.D
 Acq On : 30 Oct 2020 14:16
 Operator : SY/MD
 Sample : L4502-22ME
 Misc : 5.27g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG445ME

Quant Time: Oct 31 03:43:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 31 02:55:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	160388	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	162737	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.28	152	80315	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	163591	128.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	256.70%#
7) Chloroethane-d5	1.58	69	128089	129.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	258.72%#
11) 1,1-Dichloroethene-d2	2.10	63	254426	104.93	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	209.86%#
21) 2-Butanone-d5	3.92	46	186728	239.19	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	239.19%#
24) Chloroform-d	4.37	84	328052	138.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	276.72%#
26) 1,2-Dichloroethane-d4	5.05	65	242428	146.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	293.48%#
32) Benzene-d6	5.07	84	633759	137.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	274.56%#
36) 1,2-Dichloropropane-d6	6.09	67	192053	139.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	279.06%#
41) Toluene-d8	7.33	98	595110	134.20	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	268.40%#
43) trans-1,3-Dichloropropene-	7.64	79	110747	139.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	278.08%#
47) 2-Hexanone-d5	8.14	63	168890	283.32	ug/L	0.04
Spiked Amount	100.000	Range	45 - 130	Recovery	=	283.32%#
57) 1,1,2,2-Tetrachloroethane-	10.24	84	246222	132.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	265.34%#
64) 1,2-Dichlorobenzene-d4	11.65	152	250516	148.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	296.92%#

Target Compounds					Ovalue
42) Toluene	7.41	91	8653	1.682 ug/L	93
46) Tetrachloroethene	7.99	164	53289	48.551 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV103020\
Data File : VV019182.D
Acq On : 30 Oct 2020 14:16
Operator : SY/MD
Sample : L4502-22ME
Misc : 5.27g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG445ME

Quant Time: Oct 31 03:43:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 31 02:55:48 2020
Response via : Initial Calibration

