

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV103020\
 Data File : VV019178.D
 Acq On : 30 Oct 2020 12:38
 Operator : SY/MD
 Sample : VV1030MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK67

Quant Time: Oct 31 02:20:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 31 02:55:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	237794	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	236799	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	118977	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	72476	38.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.70%
7) Chloroethane-d5	1.58	69	59335	40.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.84%
11) 1,1-Dichloroethene-d2	2.12	63	112207	31.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.42%
21) 2-Butanone-d5	3.91	46	93676	80.93	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.93%
24) Chloroform-d	4.37	84	149376	42.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.98%
26) 1,2-Dichloroethane-d4	5.05	65	107190	43.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.52%
32) Benzene-d6	5.07	84	284501	42.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.70%
36) 1,2-Dichloropropane-d6	6.09	67	85516	42.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.40%
41) Toluene-d8	7.33	98	265731	41.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.36%
43) trans-1,3-Dichloropropene-	7.64	79	47018	40.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.14%
47) 2-Hexanone-d5	8.11	63	65652	75.69	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.69%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	107518	39.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	79.62%
64) 1,2-Dichlorobenzene-d4	11.65	152	114443	45.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.56%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV103020\
Data File : VV019178.D
Acq On : 30 Oct 2020 12:38
Operator : SY/MD
Sample : VV1030MBL01
Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK67

Quant Time: Oct 31 02:20:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 31 02:55:48 2020
Response via : Initial Calibration

