

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103023\
 Data File : VV032769.D
 Acq On : 30 Oct 2023 14:04
 Operator : SY/MD
 Sample : 05019-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 31 02:26:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 31 02:23:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	142024	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	159875	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	65229	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	41090	3.523	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	70.400%	
7) Chloroethane-d5	1.532	69	45673	4.608	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.200%	
11) 1,1-Dichloroethene-d2	2.059	65	17334	3.311	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.200%	
20) 2-Butanone-d5	0.000	46	0d	0.000	ug/L	
Spiked Amount	50.000	Range 40 - 130	Recovery	=	0.000%#	
24) Chloroform-d	4.252	84	83831	4.328	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.600%	
26) 1,2-Dichloroethane-d4	4.947	65	47472	4.965	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.200%	
32) Benzene-d6	4.966	84	149637	3.931	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	78.600%	
36) 1,2-Dichloropropane-d6	5.992	67	51852	4.704	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.000%	
41) Toluene-d8	7.246	98	129156	3.347	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	67.000%#	
43) trans-1,3-Dichloroprop...	7.561	79	14663	3.015	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	60.200%	
46) 2-Hexanone-d5	8.030	63	54258	28.149	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	56.300%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	48588	4.957	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.200%	
66) 1,2-Dichlorobenzene-d4	11.564	152	51002	4.788	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103023\
Data File : VV032769.D
Acq On : 30 Oct 2023 14:04
Operator : SY/MD
Sample : 05019-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 31 02:26:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 31 02:23:31 2023
Response via : Initial Calibration

